



AURELIA® VIBRANT

5.7g Latex Powder Free Examination Gloves

AURELIA® GLOVES

sales@supermax.co.uk
01733 615215

Supermax Healthcare Ltd
Unit 12-16 Titan Drive
Fengate, Peterborough
PE1 5XN

| | |
|----------------------|--|
| Regulations | PPE Regulation (EU) 2016/425 Medical Device Regulation (EU) 2017/745, 93/42/EEC |
| Quality Standards | ISO 9001:2015 Quality Management Systems. ISO 13485: 2016 Medical Devices Quality Management Systems. |
| Harmonised Standards | BS EN ISO 374 - 1 Protective gloves against dangerous chemicals, Type C. BS EN ISO 374 - 5 Protective gloves against microorganisms (Virus). BS EN 455 1-4 Medical gloves for single use. EN 420: 2003 + A1: 2009 Protective gloves. General requirements and test methods BS EN 1186:2002 Materials and articles in contact with foodstuffs |




10 x 100

XS S M L XL

98225 98226 98227 98228 98229

Sizes & Dimensions



XS 70+10 mm
S 80+10 mm
M 95+10 mm
L 110+10 mm
XL > 110 mm



min. 240 mm
min. 240 mm
min. 240 mm
min. 240 mm
min. 240 mm

Single Wall Thickness



① min. 0.13mm
② min. 0.10mm
③ min. 0.07mm

Physical Properties

| | Unaged | Aged |
|-------------------|---------------|---------------|
| Tensile Strength | Min. 18.0 MPa | Min. 14.0 MPa |
| Elongation | Min. 650% | Min. 500% |
| Strength at Break | Min. 6N | Min. 6N |



Latex



Powder Free



3 Years



Micro Textured



5.7g +/- 0.2g



Natural

